

# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:*

***Lab Service Tech Calibration Laboratory  
1000 Management Way  
Garner, NC 27529***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025: 2005***

*This accreditation demonstrated technical competence for the following scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated June 18, 2005):*

***Laboratory and Field Calibration of Pipettes and Balances  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

---

President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

*Initial Accreditation Date:*

June 11, 2006

*Accreditation No.:*

59382

*Issue Date:*

May 09, 2008

*Certificate No.:*

L08-26

*Expiration Date:*

May 08, 2010

*Page No.:*

Page 1 of 2

# Certificate of Accreditation: Supplement

**Lab Service Tech Calibration Laboratory**  
1000 Management Way  
Garner, NC 27529

*Accreditation is granted to this facility to perform the following calibrations:*

<b>CALIBRATION FIELD</b>	<b>MEASURED QUANTITY, INSTRUMENT OR GAUGE</b>	<b>RANGE (AND SPECIFICATION WHERE APPROPRIATE)</b>	<b>BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (<math>\pm</math>)</b>
Mass	Balances / Scales	10 mg to 400 g	0.018 mg to 0.028 g (ASTM Class 1 and Class 2 Weights)
		5 kg to 20 kg	0.028 g to 2.047 g (Class 2 and Class F Weights)
Mechanical	Pipettes	To 10 $\mu$ l	0.032 $\mu$ l
		To 100 $\mu$ l	0.021 $\mu$ l
		To 1 000 $\mu$ l	1.09 $\mu$ l
		To 5 000 $\mu$ l	1.459 $\mu$ l
		To 10 000 $\mu$ l	2.40 $\mu$ l